


Declaration of Conformity

Manufacturer Name and Address:	Cerus Corporation 2550 Stanwell Drive Concord, CA 94520 USA
EC Representative Name and Address:	Cerus Europe B.V. Stationsstraat 79-D 3811 MH Amersfoort The Netherlands
Product Name:	INTERCEPT Illuminator
Product Code or Model Number:	INT100-50 INT100-60
Global Medical Device Nomenclature (GMDN)	44185
Directive:	Council Directive 93/42/EEC
Medical Device Classification:	Class I
Conformity Assessment Route:	Annex VII and IX (Rules 1 and 12)
<p>We declare that the product(s) described above meets the applicable provisions of the above-mentioned Directive. In addition, the requirements of 2007/47/EC and 2002/96/EC (WEEE) have been fulfilled. All supporting documentation is retained under the premises of the manufacturer.</p>	
Standards Applied:	EN 1041, EN ISO 13485, EN ISO 14155-1, EN ISO 14971, EN ISO 15225, ISO 9001, ISO 7000, ISO 7064, ISO 15223-1, ISO/IEC 14882, EN 61000-3-2, EN 61000-3-3, EN 61010-1, EN 55011, EN 61000-4-1, EN 61000-4-2, EN 61000-4-3, EN 61000-4-4, EN 61000-4-5, EN 61000-4-6, EN 61000-4-8, EN 61000-4-11, EN 50419, EN 61326, IEC 62304
Certification Body (if applicable):	Not applicable
EC Certificate Number(s):	Not applicable- Self declaration
Date of CE Mark:	21 August 2006
<p>This declaration is issued under the sole responsibility of Cerus Corporation.</p>	

Declaration of Conformity

Place, Date of Issue	Concord, CA USA _____ City	August 26, 2014 _____ Date	
Name and Title of the Issuer:	Carol M. Moore _____ (Name)	Sr. VP, Regulatory Affairs, Quality and Clinical _____ (Title)	 _____ Signature